



GORE® VIABAHN® Endoprosthesis with Heparin Bioactive Surface

CONFIGURATIONS



Sizing table

TIP to HUB Device Deployment — 0.035" guidewire compatibility (radiopaque markers on 5–8 mm devices)

Device Sizing		Introducer Sheath (Fr)						Recommended Balloon Diameter for Device Touch-up ¹ (mm)
Endoprosthesis Labeled Diameter* (mm)	Recommended Vessel Diameter** (mm)	2.5 cm Device Length*	5 cm Device Length*	7.5 cm Device Length*	10 cm Device Length*	15 cm Device Length*	25 cm Device Length*	
5	4.0–4.7	7	7	7	7	7	7	5
6	4.8–5.5	7	7	7	7	7	7	6
7	5.6–6.5	8	8	8	8	8	8	7
8	6.6–7.5	8	8	9	8	8	8	8
9	7.6–8.5	–	9	9	9	9	–	9
10	8.6–9.5	11 [†]	11 [†]	–	11 [†]	11 [†]	–	10
11	9.6–10.5	11	11	–	11	–	–	12
13	10.6–12.0	12	12	–	12	–	–	14

TIP to HUB Device Deployment — 0.014" or 0.018" guidewire compatibility (with radiopaque markers)

Device Sizing		Introducer Sheath (Fr)						Recommended Balloon Diameter for Device Touch-up ¹ (mm)
Endoprosthesis Labeled Diameter* (mm)	Recommended Vessel Diameter** (mm)	2.5 cm Device Length*	5 cm Device Length*	7.5 cm Device Length*	10 cm Device Length*	15 cm Device Length*	25 cm Device Length*	
5	4.0–4.7	6	6	6	6	6	6	5
6	4.8–5.5	6	6	6	6	6	6	6
7	5.6–6.5	7	7	7	7	7	7	7
8	6.6–7.5	7	7	7	7	7	7 [¶]	8

Catalogue listings

0.035" Guidewire

(5–8 mm diameters with radiopaque markers; 9–13 mm diameters without radiopaque markers)

Catalogue Number	Endoprosthesis Labeled Diameter* (mm)	Endoprosthesis Length* (cm)	Catheter Length [¶] (cm)	Recommended Vessel Diameter** (mm)	Device Profile (Fr)
VBHR050202A	5	2.5	120	4.0–4.7	7
VBHR050502A	5	5	120	4.0–4.7	7
VBHR050702A	5	7.5	120	4.0–4.7	7
VBHR051002A	5	10	120	4.0–4.7	7
VBHR051502A	5	15	120	4.0–4.7	7
VBHR052502A	5	25	120	4.0–4.7	7
VBHR060201A	6	2.5	75	4.8–5.5	7
VBHR060202A	6	2.5	120	4.8–5.5	7
VBHR060501A	6	5	75	4.8–5.5	7
VBHR060502A	6	5	120	4.8–5.5	7
VBHR060702A	6	7.5	120	4.8–5.5	7
VBHR061001A	6	10	75	4.8–5.5	7


Catalogue Number	Endoprosthesis Labeled Diameter* (mm)	Endoprosthesis Length* (cm)	Catheter Length ^{II} (cm)	Recommended Vessel Diameter** (mm)	Device Profile (Fr)
VBHR061002A	6	10	120	4.8–5.5	7
VBHR061501A	6	15	75	4.8–5.5	7
VBHR061502A	6	15	120	4.8–5.5	7
VBHR062502A	6	25	120	4.8–5.5	7
VBHR070201A	7	2.5	75	5.6–6.5	8
VBHR070202A	7	2.5	120	5.6–6.5	8
VBHR070501A	7	5	75	5.6–6.5	8
VBHR070502A	7	5	120	5.6–6.5	8
VBHR070702A	7	7.5	120	5.6–6.5	8
VBHR071001A	7	10	75	5.6–6.5	8
VBHR071002A	7	10	120	5.6–6.5	8
VBHR071501A	7	15	75	5.6–6.5	8
VBHR071502A	7	15	120	5.6–6.5	8
VBHR072502A	7	25	120	5.6–6.5	8
VBHR080201A	8	2.5	75	6.6–7.5	8
VBHR080202A	8	2.5	120	6.6–7.5	8
VBHR080501A	8	5	75	6.6–7.5	8
VBHR080502A	8	5	120	6.6–7.5	8
VBHR080702A	8	7.5	120	6.6–7.5	8
VBHR081001A	8	10	75	6.6–7.5	8
VBHR081002A	8	10	120	6.6–7.5	8
VBHR081501A	8	15	75	6.6–7.5	8
VBHR081502A	8	15	120	6.6–7.5	8
VBHR082502A	8	25	120	6.6–7.5	8
VBH090502A	9	5	120	7.6–8.5	9
VBH090702A	9	7.5	120	7.6–8.5	9
VBH091002A	9	10	120	7.6–8.5	9
VBH091502A	9	15	120	7.6–8.5	9
VBH100202A	10	2.5	120	8.6–9.5	11 [†]
VBH100502A	10	5	120	8.6–9.5	11 [†]
VBH101002A	10	10	120	8.6–9.5	11 [†]
VBH101502A	10	15	120	8.6–9.5	11 [†]
VBH110202A	11	2.5	120	9.6–10.5	11
VBH110502A	11	5	120	9.6–10.5	11
VBH111002A	11	10	120	9.6–10.5	11
VBH130202A	13	2.5	120	10.6–12.0	12
VBH130502A	13	5	120	10.6–12.0	12
VBH131002A	13	10	120	10.6–12.0	12

Catalogue listings

0.014 / 0.018" Guidewire

(5–8 mm diameters with radiopaque markers)

Catalogue Number	Endoprosthesis Labeled Diameter* (mm)	Endoprosthesis Length* (cm)	Catheter Length ^{II} (cm)	Recommended Vessel Diameter** (mm)	Device Profile (Fr)
VBJR050202A	5	2.5	120	4.0–4.7	6
VBJR050502A	5	5	120	4.0–4.7	6
VBJR050702A	5	7.5	120	4.0–4.7	6
VBJR051002A	5	10	120	4.0–4.7	6
VBJR051502A	5	15	120	4.0–4.7	6
VBJR052502A	5	25	120	4.0–4.7	6
VBJR060202A	6	2.5	120	4.8–5.5	6
VBJR060502A	6	5	120	4.8–5.5	6
VBJR060702A	6	7.5	120	4.8–5.5	6
VBJR061002A	6	10	120	4.8–5.5	6
VBJR061502A	6	15	120	4.8–5.5	6
VBJR062502A	6	25	120	4.8–5.5	6
VBJR070202A	7	2.5	120	5.6–6.5	7
VBJR070502A	7	5	120	5.6–6.5	7
VBJR070702A	7	7.5	120	5.6–6.5	7
VBJR071002A	7	10	120	5.6–6.5	7
VBJR071502A	7	15	120	5.6–6.5	7
VBJR072502A	7	25	120	5.6–6.5	7
VBJR080202A	8	2.5	120	6.6–7.5	7
VBJR080502A	8	5	120	6.6–7.5	7
VBJR080702A	8	7.5	120	6.6–7.5	7
VBJR081002A	8	10	120	6.6–7.5	7
VBJR081502A	8	15	120	6.6–7.5	7
VBJR082502A	8	25	120	6.6–7.5	7 [¶]

 Consult Instructions for Use

INTENDED USE / INDICATIONS: The GORE® VIABAHN® Endoprosthesis is indicated for improving blood flow in patients with symptomatic peripheral arterial disease in superficial femoral artery de novo and restenotic lesions up to 270 mm in length with reference vessel diameters ranging from 4.0 – 7.5 mm, in superficial femoral artery in-stent restenotic lesions up to 270 mm in length with reference vessel diameters ranging from 4.0 – 6.5 mm, and in iliac artery lesions up to 80 mm in length with reference vessel diameters ranging from 4.0 – 12 mm. The GORE® VIABAHN® Endoprosthesis is also indicated for the treatment of stenosis or thrombotic occlusion at the venous anastomosis of synthetic arteriovenous (AV) access grafts. Refer to Instructions for Use at goremedical.com for a complete description of all contraindications, warnings, precautions and adverse events. ^{Rx only}



W. L. GORE & ASSOCIATES, INC.
Flagstaff, AZ 86004

+65.67332882 (Asia Pacific) 800.437.8181 (United States)
00800.6334.4673 (Europe) 928.779.2771 (United States)

goremedical.com

Products listed may not be available in all markets.

SUPER SHEATH is a trademark of Boston Scientific. INTRADYN is a trademark of B. Braun Medical, Inc. COOK® and CHECK-FLO® are trademarks of Cook Medical, Inc. AVANTI® is a trademark of Cordis Corporation.

GORE®, TIP to HUB, VIABAHN®, and designs are trademarks of W. L. Gore & Associates.
© 2007–2014, 2016, 2018 W. L. Gore & Associates, Inc. AL0343-EN13 SEPTEMBER 2018

* Labeled device diameters and lengths are nominal.

** Recommended endoprosthesis compression within the vessel is approximately 5 – 20%.

† For the 11 and 13 mm diameter devices, balloon inflation pressure should not exceed 8 atm.

‡ The 10 mm diameter device is compatible with the following 10 Fr introducer sheaths:
Cordis AVANTI® Sheath Introducer, Boston Scientific SUPER SHEATH® Introducer Sheath, B. Braun INTRADYN Tear-Away Introducer Sheath.

II Ensure the guidewire is the appropriate size (see *Instructions for Use*) and has a length at least twice that of the delivery catheter.

¶ The 8 mm x 25 cm device is not compatible with the 7 Fr COOK® CHECK-FLO® Sheath