

GORE® VIABAHN®

Endoprosthesis with
PROPATEN Bioactive Surface

Together, improving life



GORE® VIABAHN®
ENDOPROSTHESIS WITH
PROPATEN BIOACTIVE SURFACE*
— CONFIGURATIONS

* As used by Gore, PROPATEN Bioactive Surface refers to Gore's proprietary CBAS® Heparin Surface.



Sizing Table

0.035" Guidewire compatible

Device Sizing		Introducer Sheath (Fr)						
Endoprosthesis labeled diameter (mm)*	Recommended vessel diameter† (mm)	2.5 cm Device Length*	5 cm Device Length*	7.5 cm Device Length*	10 cm Device Length*	15 cm Device Length*	25 cm Device Length*	Recommended Balloon Diameter for Device Touch-up† (mm)
5	4.0–4.7	7	7	7	7	7	7	5
6	4.8–5.5	7	7	7	7	7	7	6
7	5.6–6.5	8	8	8	8	8	8	7
8	6.6–7.5	8	8	8	8	8	8	8
9	7.6–8.5	–	8	8	8	8	–	9
10	8.6–9.5	–	8	–	8	8	–	10
11	9.6–10.5	–	10	–	10	–	–	12
13	10.6–12.0	–	10 ^s	–	10 ^s	–	–	14

0.018" Guidewire compatible

Device Sizing		Introducer Sheath (Fr)						
Endoprosthesis labeled diameter (mm)*	Recommended vessel diameter† (mm)	2.5 cm Device Length*	5 cm Device Length*	7.5 cm Device Length*	10 cm Device Length*	15 cm Device Length*	25 cm Device Length*	Recommended Balloon Diameter for Device Touch-up (mm)
5	4.0–4.7	6	6	6	6	6	6	5
6	4.8–5.5	6	6	6	6	6	6	6
7	5.6–6.5	7	7	7	7	7	7	7
8	6.6–7.5	7	7	7	7	7	7 ^{II}	8

Catalogue listings

0.018" Guidewire compatible


Catalogue Number	Endoprosthesis Labeled Diameter* (mm)	Endoprosthesis Length* (cm)	Catheter Length† (cm)	Recommended Vessel Diameter† (mm)	Device Profile (Fr)
PAJR050202E	5	2.5	120	4.0–4.7	6
PAJR050502E	5	5	120	4.0–4.7	6
PAJR050702E	5	7.5	120	4.0–4.7	6
PAJR051002E	5	10	120	4.0–4.7	6
PAJR051502E	5	15	120	4.0–4.7	6
PAJR052502E	5	25	120	4.0–4.7	6
PAJR060202E	6	2.5	120	4.8–5.5	6
PAJR060502E	6	5	120	4.8–5.5	6
PAJR060702E	6	7.5	120	4.8–5.5	6
PAJR061002E	6	10	120	4.8–5.5	6
PAJR061502E	6	15	120	4.8–5.5	6
PAJR062502E	6	25	120	4.8–5.5	6
PAJR070202E	7	2.5	120	5.6–6.5	7
PAJR070502E	7	5	120	5.6–6.5	7
PAJR070702E	7	7.5	120	5.6–6.5	7
PAJR071002E	7	10	120	5.6–6.5	7
PAJR071502E	7	15	120	5.6–6.5	7
PAJR072502E	7	25	120	5.6–6.5	7
PAJR080202E	8	2.5	120	6.6–7.5	7
PAJR080502E	8	5	120	6.6–7.5	7
PAJR080702E	8	7.5	120	6.6–7.5	7
PAJR081002E	8	10	120	6.6–7.5	7
PAJR081502E	8	15	120	6.6–7.5	7
PAJR082502E	8	25	120	6.6–7.5	7 ^{II}

Sizing Table

0.035" Guidewire compatible

Catalogue Number	Endoprosthesis Labeled Diameter* (mm)	Endoprosthesis Length* (cm)	Catheter Length† (cm)	Recommended Vessel Diameter† (mm)	Device Profile (Fr)
PAHR050202E	5	2.5	120	4.0–4.7	7
PAHR050502E	5	5	120	4.0–4.7	7
PAHR050702E	5	7.5	120	4.0–4.7	7
PAHR051002E	5	10	120	4.0–4.7	7
PAHR051502E	5	15	120	4.0–4.7	7
PAHR052502E	5	25	120	4.0–4.7	7
PAHR060201E	6	2.5	75	4.8–5.5	7
PAHR060202E	6	2.5	120	4.8–5.5	7
PAHR060501E	6	5	75	4.8–5.5	7
PAHR060502E	6	5	120	4.8–5.5	7
PAHR060702E	6	7.5	120	4.8–5.5	7
PAHR061001E	6	10	75	4.8–5.5	7
PAHR061002E	6	10	120	4.8–5.5	7
PAHR061501E	6	15	75	4.8–5.5	7
PAHR061502E	6	15	120	4.8–5.5	7
PAHR062502E	6	25	120	4.8–5.5	7
PAHR070201E	7	2.5	75	5.6–6.5	8
PAHR070202E	7	2.5	120	5.6–6.5	8
PAHR070501E	7	5	75	5.6–6.5	8
PAHR070502E	7	5	120	5.6–6.5	8
PAHR070702E	7	7.5	120	5.6–6.5	8
PAHR071001E	7	10	75	5.6–6.5	8
PAHR071002E	7	10	120	5.6–6.5	8
PAHR071501E	7	15	75	5.6–6.5	8
PAHR071502E	7	15	120	5.6–6.5	8
PAHR072502E	7	25	120	5.6–6.5	8
PAHR080201E	8	2.5	75	6.6–7.5	8
PAHR080202E	8	2.5	120	6.6–7.5	8
PAHR080501E	8	5	75	6.6–7.5	8
PAHR080502E	8	5	120	6.6–7.5	8
PAHR080702E	8	7.5	120	6.6–7.5	8
PAHR081001E	8	10	75	6.6–7.5	8
PAHR081002E	8	10	120	6.6–7.5	8
PAHR081501E	8	15	75	6.6–7.5	8
PAHR081502E	8	15	120	6.6–7.5	8
PAHR082502E	8	25	120	6.6–7.5	8
PAHR090502E	9	5	120	7.6–8.5	8
PAHR090702E	9	7.5	120	7.6–8.5	8
PAHR091002E	9	10	120	7.6–8.5	8
PAHR091502E	9	15	120	7.6–8.5	8
PAHR100502E	10	5	120	8.6–9.5	8
PAHR101002E	10	10	120	8.6–9.5	8
PAHR101502E	10	15	120	8.6–9.5	8
PAHR110502E	11	5	120	9.6–10.5	10
PAHR111002E	11	10	120	9.6–10.5	10
PAHR130502E	13	5	120	10.6–12.0	10 ⁶
PAHR131002E	13	10	120	10.6–12.0	10 ⁶

KEY: PRODUCT NUMBERING SYSTEM (EXAMPLE ONLY)



PAH Heparin 0.035" Guidewire Compatibility

PAJ Heparin 0.018" Guidewire Compatibility

* Labeled device diameters and lengths are nominal.

† Recommended endoprosthesis compression within the vessel is approximately 5 – 20%.

‡ For the 11 and 13 mm diameter devices, balloon inflation pressure should not exceed 8 atm.

§ The 13 mm diameter device is not compatible with the 10 Fr COOK® FLEXOR® CHECK-FLO® Introducer.

|| The 8 mm x 25 cm device is not compatible with the 7 Fr COOK® FLEXOR® CHECK-FLO® Introducer.

¶ Ensure the guidewire is the appropriate size (see *Instructions for Use*) and has a length at least twice that of the delivery catheter.

 Consult Instructions
for Use
eifu.goremedical.com

Refer to *Instructions for Use* at eifu.goremedical.com for a complete description of all applicable indications, warnings, precautions and contraindications for the markets where this product is available. Rx Only

Products listed may not be available in all markets.

COOK, CHECK-FLO and FLEXOR are trademarks of Cook Medical, Inc.

CBAS is a trademark of Carmeda AB, a wholly owned subsidiary of W. L. Gore & Associates, Inc.

GORE, *Together, improving life*, PROPATEN, VIABAHN and designs are trademarks of W. L. Gore & Associates.

© 2024 W. L. Gore & Associates GmbH 24PL2177-EN01 DECEMBER 2024

W. L. Gore & Associates, Inc.
goremedical.com

Asia Pacific +65 6733 2882 **Australia/New Zealand** 1800 680 424 **Europe** 00800 6334 4673
United States Flagstaff, AZ 86003 800 437 8181 928 779 2771

